K093697

Exhibit 3: 510(k) Summary VASHE™ WOUND THERAPY SOLUTION

510 (k) Summary	This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of 21 C.F.R. §807.92.
Submitter	PuriCore Inc. 508 Lapp Road Malvern, PA 19355
Contact Person	Dennis Mahoney PuriCore Inc. 508 Lapp Road Malvern, PA 19355 484-321-2724 (ph) ; 610-341-0503 (fax) APR 13 2010
Date Prepared	November 20th , 2009
Trade Name	Vashe® Wound Therapy Solution
Common Name	Wound Cleanser
Classification Name	Solution, saline, (wound dressing)
Predicate Device	Anasept™ Antimicrobial Skin and Wound Cleanser; Anacapa™ Technologies, Inc. K073547, April 23rd, 2008, Oculus Puracyn™ Skin and Wound Cleanser with Preservatives, Oculus Innovative Sciences, K090206, June 2nd, 2009, Microcyn™ Skin Wound Gel, Oculus Innovative Sciences, K090725, May 20th 2009, Dermacyn™ Wound Cleanser and Wound Dressing, K041161, May 3rd, 2005.
Description	The subject device is a wound cleansing solution that is intended for cleansing, irrigating, debriding dermal wounds in addition to moistening and lubricating absorbent wound dressings. The mechanical action of fluid moving across the wound provides for the mechanism of action and aids in the removal of foreign objects such as dirt and debris.
Indications for Use	Vashe® Wound Therapy Solution is intended for OTC use for management of minor skin abrasions, minor lacerations, minor irritations, minor cuts, and intact skin.
Substantial Equivalence	The product is similar in function and intended use to: • Anasept™ Antimicrobial Skin and Wound Cleanser manufactured by Anacapa™ Technologies, Inc. is intended for OTC use for management of skin abrasions, lacerations, minor irritations, cuts, exit sites and intact skin. • Oculus Puracyn™ Skin and Wound Cleanser with Preservative manufactured by Oculus Innovative Sciences, is intended for OTC use for management of skin abrasions, lacerations, minor irritations, cuts, exit sites and intact skin. • Microcyn™ Skin Wound Gel by Oculus Innovative Sciences is intended for the management of minor abrasions, lacerations, cuts, and intact skin. • Dermacyn™ Wound Cleanser and Wound Dressing manufactured by Oculus Innovative Sciences, is intended for OTC use for management of skin abrasions, lacerations, minor irritations, cuts, exit sites and intact skin.
Non-clinical Performance	Pre-clinical testing demonstrated biocompatibility of the Vashe® Wound Therapy Solution.
Conclusion	Vashe® Wound Therapy Solution is substantially equivalent to the currently cleared and marketed Anasept™ Antimicrobial Skin and Wound Cleanser, Oculus Puracyn™ Skin and Wound Cleanser with Preservative, Microcyn™ Skin Wound Gel, and Dermacyn™ Wound Cleanser and Wound Dressing.

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room –WO66-G609
Silver Spring, MD 20993-0002

PuriCore, Inc.
% Mr. Dennis Mahoney
Director of Quality Assurance &
Regulatory Affairs
508 Lapp Road
Malvern, Pennsylvania 19355

APR 13 2010

Re: K093697
Trade/Device Name: Vashe® Wound Therapy Solution
Regulatory Class: Unclassified
Product Code: FRO
Dated: April 1, 2010
Received: April 2, 2010

Dear Mr. Mahoney:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,



Mark N. Melkerson
Director
Division of Surgical, Orthopedic
and Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Exhibit 2

Indications for Use Statement

510(k) Number:

Device Name : Vashe® Wound Therapy Solution

Indications for Use :

Vashe® Wound Therapy System is intended for OTC use for management of minor skin abrasions, minor lacerations, minor irritations, minor cuts, and intact skin.

Prescription Use _____ OR Over-The-Counter Use: XXX
(Per 21 CFR 801.109)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED

Concurrence of CDHR, Office of Device Evaluation (ODE)

David Krause for MXM
(Division Sign-Off)
Division of Surgical, Orthopedic, and Restorative Devices

510(k) Number K093697